How to fill out for an order form of a wire additional order

About this order form

This order form can be used when you want to corrections to torque, angulation, etc during the treatment period.

Please select the wire according to doctor's treatment plan and fill out the correction values.

How to about order

■ Please fill out one wire per piece of order form due to the instructions for wire correction values are complicated.

However, you can use one order form if there are multiple lines with the same correction values or if there is one line on the top and one line on the bottom even if the corrections are different.

■ Please fill out all required information in English.

Please fill out the doctor's name and contact information as they are all required.

■ Please fill out with black or blue pen.

Also, please do not use pencil, erasable pen and correction tape.

■ We will proceed with constructing products based on the data you provided with the first order.

You don't need to send impressions and models unless there is an additional change in a treatment plan.

■ Please send an order form you filled out to Aso International, Inc.

ASO International, Inc.

TEL: 03-3547-0471 / FAX: 03-3547-0475

Email: lingual@aso-inter.co.jp

■ This order form data can be download from the 3M website below

ASO Web Site

URL: https://www.aso-inter.co.jp/download/

*This website doesn't require logging in for downloading.

Order Form for Spare Parts [Incognito Appliance System]

ASO International, Inc. ASO International, And Orthodontic Laboratory Services

FAX:03-3547-0475

Please fill out all required fields in English with black or blue pen.

E-mail:lingual@aso-inter.co.jp

Do n :Attent	correction	ce las ecificat on value e corre	ions are	e crown e compl llues are	related	d! Torqu	e use c nultiple	one inst	ruction	sheet p	er page	 fer but	only on	i-Wires! e wire is		d for b	oth the upper ar	and lo	wer teeth, you ma Set of archv		
oth e lation mesial	tion: Sp	ecificat on value e corre	es can b ction va	e crown e compl llues are	n related, licated, the sai	d! Torqu so pleas me for n	e use c nultiple	one inst	ruction	sheet p	er page tions dif	fer but	only on	e wire is		d for b	oth the upper ar	nd lo			
lation lon mesial	8							1	1		. Qua	dran	. / L ºº						Set of archy	vires M	
lation lon mesial	8	7	6	5	4	3	2	1	1)		Crown movement		Wires required	straight	
ion mesial distal										2	3	4	5	6	7	8	歯冠の動き + = 頬側方向 - = 舌側方向	_	(please tick) 12 Ø 14 Ø		lat. se
mesial distal																	+ = 近心側 - = 遠心側	SF NiTi	16 Ø	E	
distal							5										+ = 近心方向 - = 遠心方向 + = 頬側方向		17 x 25 18 x 25	E	
al																	-= 舌側方向 += 頬側方向	<u> </u>	16 x 22 16 x 22 ET* 16 x 24	4	1
																	-= 舌側方向 += 頬側方向 -= 舌側方向	Ster	16 x 24 16 x 24 ET** 18 x 25	E	
sal							?	\									+ = 挺出 - = 圧下	anii d	18 x 25 (red.) 17,5 x 17,5 17 x 25		
ger 長 ment 特	特記事項		horter	短め		distan	7	ween ———	;	and			mm					Beta III Titapium	18,2 x 18,2 18,2 x 25		Е
= 15度	ミトルクス	.り(部位	:上顎両(則中切歯) ET**	= 13度	トルク入	.り(部位	:上顎6前	j歯) re	ed. = 側	方歯群り	Jデュ ー ス	. Ε = <u>‡</u>	支歯症例	世奨ワイ	´ヤー N = 非抜b	歯症 の	列推奨ワイヤー ■	■= 取り扱	 及い無
oth	8	4	. Qu a	idrant	t (下顎 4	頁右側) 3	2	1	1	2	3. Qu a	dran 4	t (下号	頁左側 6	7	8	Crown movement 歯冠の動き		Set of archy Wires required (please tick)	vires M straight lat. sec.	
e		,	0		7			•	'		0		0		,	0	+ = 頬側方向 - = 舌側方向		12 Ø 14 Ø	E	lat
lation																	+ = 近心側 - = 遠心側	SENIT		E	1
ion mesial							<u> </u>)									- = 遠心方向 + = 頬側方向		18 x 25	E	1
distal																	- = 古側方向 - = 括側方向	Steel		4	١
al																	+ = 類側方向 - = 舌側方向	anii ac			
sal	.b	S	horter	短め「	<u> </u>]	distan	6 t	ween		and]mm				+ = 挺出 - = 圧下	Rota III Tit	18,2 x 18,2 18,2 x 25		EN
ger 長						(7)	<u> </u>				1								
lat ior me	ion sial ttal	ion isial ital i	ion isial tal	ion isial ital i er 長め Shorter	ion isial ital i er 長め Shorter 短め	ion isial ital i er 長め Shorter 短め	ion isial ital i er 長め	ion sial tal er 長め Shorter 短め distar	ion	ion isial ital ier 長め Shorter 短め distanctitiveen :	ion	ion isial ital ier 長め Shorter 短め distanctitiveen and [ion	ion	ion	ion	ion	+ = 類側方向	H = 類側方向	H = 類側方向	12 Ø 14 Ø E 16 Ø 16 Ø Ø 16 Ø Ø Ø Ø Ø Ø Ø Ø Ø Ø Ø Ø Ø Ø Ø Ø Ø Ø

Patient and Clinic Information

Patient name	Kyousei Hanako	Orthodontist	Ginza Taro
Orthodontic Practice	ASO Orthodontic Office	Phone 81-3-3	547-0471

- Please ensure that both the doctor's and patient's full names are entered.
- Please provide the clinic's name and phone number accurately.

2. About Adjustments

X Do not take last correction values.

If you do not wish to include the adjustments made to your previous order, please check X this box. If this box is left unchecked, the adjustments from your previous order will be included.

If you would like to enter ideal values based on your initial consultation, please check X this box.

3. Notes on ordering additional wires with correction values

NOTE 1) Attention: Specifications are crown related! Torque, Angulation, Occlusal is not possible on SE NiTi-Wires!

NOTE 2) These instructions are for one wire only.

However, you can use one sheet for multiple wires with the same adjustments, or one wire per upperand lower wire even if the adjustments are different.

Regarding NOTE 1

SE NiTi wires cannot be modified for torque, angulation, or occlusal adjustments.

Please enter any requests other than these three parameters.

Regarding NOTE 2

Because specifying wire adjustment values can be complex, we ask that you submit **one** instruction sheet per wire.

However, you may use a single sheet for multiple wires if the adjustment values are the same, or for one wire each in the upper and lower arches—even if the adjustments differ.

4. Wire Sequence

Se	t of archwir	es MA	(
	res required ease tick)	straight lat. sec.	
	12 Ø		
	14 Ø		N
Ē	16 Ø	Е	
SE NITI	16 x 22	Е	Ν
	17 x 25		
	18 x 25	Е	Ν
	16 x 22		
	16 x 22 ET*		
Steel	16 x 24		N
Ste	16 x 24 ET**	ΕX	
	18 x 25		
	18 x 25 (red.)		
E	17,5 x 17,5		
Beta III Titanium	17 x 25		
=	18,2 x 18,2		E N
Bet	18,2 x 25		

Se	t of archwir	es MDI	3
	res required ease tick)	straight lat. sec.	indiv. lat. sec.
	12 Ø		
	14 Ø	Е	Ν
Ë	16 Ø		
SE NiTi	16 x 22	Е	N
	17 x 25		
	18 x 25	Е	N
	16 x 22		
Steel	16 x 24	Е	N
Ste	18 x 25		
	18 x 25 (red.)		
E	17,5 x 17,5		X
itani	17 x 25		
Beta III Titanium	18,2 x 18,2		E N
Beta	18,2 x 25		

Please select a wire based on your own treatment plan.

Wire Types and Notations

- straight lat. sec: A wire with a straight molar area. Primarily for extraction cases.
- indv. lat. sec.: An individual wire with bends throughout.

 Primarily for non-extraction cases.
- ET*: 15° additional torque on maxillary central incisors
- ET^{**} : 30° additional torque on the 6 upper anterior teeth
- red.: Reduced wire for lateral teeth

 The lateral teeth area has a reduced wire to facilitate easier sliding.
- **Straight Beta IIITitanium wires are not available**; please use the individual wires instead.
- *Wires are usually selected based on your treatment plan. If you do not have a preference, please select the recommended E or N wires, or refer to the specific wire sequence for each case.

5. Modification Details

Please use + and to enter numbers to indicate how you would like the **crown** to move.

- * Please note that root movement cannot be recorded.
- * SE NiTi does not allow for torque, angulation, or occlusal corrections.

			1	. Qua	adran	t (上咢	質右側)			2	. Qua	ıdran	t (上号	質左側)		Crown movement
	Tooth	8	7	6	5	4	3	2	1	1	2	3	4	5	6	7	8	歯冠の動き
角度	Torque																	+ = 頬側方向 = 舌側方向
degrees 角	Angulation								A									+ = 近心側 = 遠心側
ep	Rotation																	+ = 頬側方向 = 舌側方向
	Only mesial edge																	+ = 頬側方向 = 舌側方向
移動量	Only distal edge																	+ = 頬側方向 = 舌側方向
mm *	Buccal								L)								+ = 頬側方向 = 舌側方向
	Occlusal																	+ = 挺出 = 圧下

A: Regarding Torque, Angulation, and Rotation:

Please enter the angle to which you would like the crown to move.

* Torque and Angulation cannot be set for SE NiTi.

角度	Torque		+3°	+2°							+ = 頬側方向 = 舌側方向
degrees 角	Angulation										+ = 近心側 = 遠心側
ge	Rotation						-2°				+ = 頬側方向 = 舌側方向

B: Regarding Only mesial edge, Only distal edge, Buccal, and Occlusal:

Please indicate the amount of crown movement you would like to make.

* SE NiTi cannot be used for occlusal placement.

	Only mesial edge			-1.6							+ = 頬側方向 = 舌側方向
移動量	Only distal edge									+2	+ = 頬側方向 = 舌側方向
mm &	Buccal										+ = 頬側方向 = 舌側方向
	Occlusal					+0.5					+ = 挺出 = 圧下

- · Only mesial edge: Moves only the mesial side buccolingually, with the distal side as the axis.
- · Only distal edge: Moves only the distal side buccolingually, with the mesial side as the axis.
- · Buccal: Moves the crown buccolingually.
- · Occlusal: Extrusion and intrusion.

6. Wire Length

Longer 長め 💢	Shorter 短め	distance between 3 and 3 2 mm
Longo Ko	Chorton /#es	

You can customize the wire length.

Please specify how much longer or shorter you want the wire to be between each tooth.

7. Special Notes

Comment (Upper):		
Comment (Opper).		

If there are any special notes, please write them in English.

8. Signature Information

Since the product will be produced overseas, your handwritten date and signature are required Please confirm your order details and fill out at the end.